

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**LESAK s.r.o.**  
CAB number 2411, Calibration Laboratory  
Božetěchova 2826/36, Královo Pole, 612 00 Brno

**CMC for the field of measured quantity: Mass**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min	unit	max	unit					
1*	Scales with non-automatic function			0 g		class E2	0.006 mg	Loading using a reference weight (according to OIML R111-1:2004)	KP 1	
		1 mg	to	200 mg			0.018 mg			
		200 mg	to	1 g			0.026 mg			
		1 g	to	2 g			0.032 mg			
		2 g	to	10 g			0.050 mg			
		10 g	to	20 g			0.064 mg			
		20 g	to	50 g			0.078 mg			
		50 g	to	100 g			0.13 mg			
		100 g	to	6 kg		class F1	1.3.10 <sup>-6</sup>			
		6 kg	to	30 kg		class F2	6.10 <sup>-6</sup>			
		30 kg	to	70 kg		class M1	2.2.10 <sup>-5</sup>			
		70 kg	to	10 t			7.3.10 <sup>-5</sup>			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

*"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself. "*