

**The Appendix is an integral part of
Certificate of Accreditation No. 238/2024 of 28/05/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

EUCAL s.r.o.
CAB number 2404, Calibration Laboratory
Voděradská 2552/16, 251 01 Říčany

CMC for the field of measured quantity: Volume

Ord. number ₁	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the meas. quantity	Lowest stated expanded measurement uncertainty ^{2,4}	Calibration principle	Calibration procedure identification ³	Location
		min. unit	max. unit					
1	Mechanical and electronic piston pipettes	0.1 µl	to 0.5 µl		0.03 µl	Gravimetric comparison with a standard (distilled water)	SOP.KL ECS 11 (ČSN EN ISO 8655-6, EURAMET cg-19)	
	pipettes, dispensers, burettes, dilutors	0.5 µl 10,000 µl	to 10,000 µl to 100,000 µl		0.15 % + 0.15 µl 0.06 %			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

⁴ The lowest uncertainty includes the effect of the operator.