

**The Appendix is an integral part of
Certificate of Accreditation No: 251/2024 of 30/05/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Univerzita Palackého v Olomouci
CAB number 1308, Laboratory of Cell Cultures, Faculty of Medicine
Hněvotínská 3, 775 15 Olomouc

Detailed information on activities within the scope of accreditation (tested subject) is given in the section „Specification of the scope of accreditation“.

Tests:

Ordinal number ¹	Test procedure / method name	Test procedure / method identification ²	Tested subject	Degrees of freedom ³
1	Test for <i>in vitro</i> cytotoxicity - Test of direct contact	SOP 1 (ČSN EN ISO 10993-5, cl. 1-3, 4.1, 4.3, 4.4, 5-7, 8, 8.3, 9,10; ČSN EN ISO 10993-12; ČSN EN ISO 7405, cl. 6.2.7, 6.2.8)	Insoluble solid samples with at least one flat surface	-
2	Test for <i>in vitro</i> cytotoxicity - Test of extract	SOP 2 (ČSN EN ISO 10993-5, cl. 1-3, 4.1, 4.2, 4.4, 5-7, 8, 8.2.1-8.2.8, 8.5, 9, 10; ČSN EN ISO 10993-12)	Insoluble solid samples	-
3	Test for <i>in vitro</i> cytotoxicity - Test of soluble samples	SOP 3 (ČSN EN ISO 10993-5, cl. 1-3, 4.3.3, 4.4, 5-7, 8, 8.2.1-8.2.8, 8.5, 9, 10; ČSN EN ISO 10993-12)	Soluble solid samples, liquid samples	-

¹ asterisk at the ordinal number identifies the tests, which the laboratory is qualified to carry out outside the permanent laboratory premises

² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)

³ the laboratory does not apply a flexible approach to the scope of accreditation

Specification of the scope of accreditation:

Ordinal test number	Detailed information on activities within the scope of accreditation (tested subject)
1 - 3	Samples are medical devices defined in ČSN EN ISO 10993-1:2021

Explanation

SOP- Standard Operation Procedure

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself. "