

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Length

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1	Steel parallels	0.5 mm	to	1,000 mm		(2L + 0.2) μm	Comparative measurement using steel parallels	KP D1		
2*	Steel length gauges	0 m	to	2 m		60 μm	Comparative measurement using steel parallels	KP D2		
		2 m	to	5 m		180 μm				
	Steel tape measures	0 m	to	2 m		0.14 mm	Comparative measurement on a reference track			
		2 m	to	3 m		0.28 mm				
		3 m	to	5 m		0.42 mm				
	5 m	to	8 m		0.70 mm					
		8 m	to	10 m		0.98 mm				
3	Tape measures	0 m	to	10 m		0.4 mm	Comparative measurement on a reference track	KP D3		
		10 m	to	20 m		0.6 mm				
		20 m	to	50 m		1.0 mm				
		50 m	to	100 m		2.2 mm				
	Laser distance meters	0 m	to	5 m		0.2 mm				
		5 m	to	10 m		0,4 mm				
4	Limit and end measuring rings	1 mm	to	100 mm		(2L + 0.5) μm	Direct and comparative measurement by a distance meter	KP D4		
		100 mm	to	500 mm		(2L + 2.4) μm				
	Limit snap gauges	1 mm	to	100 mm		(2L + 0.5) μm				
		100 mm	to	500 mm		(2L + 2.4) μm				
	Feeler gauges	0.02 mm	to	100 mm		(2L + 0.5) μm				
	Limit cylindrical gauges	100 mm	to	500 mm		(2L + 2.4) μm				

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
5*	Feeler gauges, Limit cylindrical gauges.	1 mm	to	125 mm		(2L + 2.4) μm	Direct measurement by a micropasameter	KP D4		
6	Limit plug gauges	1 mm	to	200 mm		(3L + 3) μm	Direct measurement by a distance meter	KP D5		
		1 mm	to	160 mm		(1L + 4) μm	Direct measurement on MasterScanner XP 16060			
	Threaded rings	1 mm	to	3 mm		(3L + 3) μm	Comparison by a wear pin gauge			
		2.5 mm	to	200 mm		(3L + 3) μm	Indirect measurement by a distance meter			
		3 mm	to	160 mm		(1L + 4) μm	Direct measurement on MasterScanner XP 16060			
7*	Limit plug gauges	1 mm	to	125 mm		(3L + 3.5) μm	Direct measurement by a micropasameter	KP D5		
8*	Slide gauges: slide rules, depth gauges, height gauges	0 mm	to	1,000 mm		12 μm	Comparative measurement using steel parallels and ring	KP D6		
		1,000 mm	to	3,000 mm		20 μm				
9*	Micrometer gauges: micrometers, pasameters, micropasameters, micrometer heads, micrometer depth gauges	0 mm	to	25 mm		0.7 μm	Comparative measurement using steel parallels	KP D7		
		25 mm	to	100 mm		1.4 μm				
		100 mm	to	1,000 mm		2.5 μm				
		1,000 mm	to	1,500 mm		4.1 μm				
10*	Inside micrometers Three contact internal gauges	2 mm	to	100 mm		2.0 μm	Comparative measurement using setting rings	KP D8		
		100 mm	to	300 mm		4.0 μm				

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
11	Inside micrometer gauges	10 mm	to	3,000 mm		(3L + 2.2) μm	Direct measurement by a distance meter	KP D9		
12*	Electromagnetic, ultrasonic thickness gauges	0 mm	to	1.5 mm		(1·L + 1.3) μm	Comparative measurement by a thickness reference standard	KP D10		
		1.5 mm	to	500 mm		(1·L + 2.3) μm				
13	Direct and lever dial indicators	0 mm	to	100 mm		0.3 μm	Direct measurement by a special measuring device	KP D11		
	Two-contact internal gauges	2 mm	to	205 mm		0.3 μm				
14	Gauges, measuring jigs, templates, meters of plane and angle	0 mm	to	2,000 mm		(4.5L + 1.7) μm	Measurement by a 3D CMM	KP D12		
15*	Profile projectors, measuring microscopes	0 mm	to	300 mm		(1·L + 2.6) μm	Comparative measurement using a rule	KP D13		
16*	Measurement of straightness, linear sensing, measurement flatness engineering gauges	0 m	to	20 m		(1·L + 0.1) μm	Direct measurement by a laser interferometer	KP D14		
		0 m	to	20 m		1.5 μm/m ²				
17	Gauges, measuring jigs, special meters	0 mm	to	600 mm		(2.5L + 1.2) μm	Measurement by a linear height gauge	KP D15		
18*	Linear height gauges	0 mm	to	600 mm		(0.8L + 0.5) μm	Comparative measurement by a calibration comb	KP D16		
		600 mm	to	1,000 mm		(1·L + 3.0) μm	Comparative measurement by a calibration comb and steel parallels			
19*	Contourographs	0 mm	to	100 mm		(1·L + 2.6) μm	Comparative measurement using end standards	KP D17		

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
20*	Length measuring instruments	0 mm	to	20,000 mm		(2L + 0.2) μm	Direct measurement by a laser interferometer	KP D18		
21*	3D coordinate measuring machines	0 mm	to	600 mm		(2L + 0.2) μm	Comparative measurement by a calibration comb	KP D19		
		600 mm	to	1,000 mm		(2L + 0.2) μm	Comparative measurement by a calibration comb and steel parallels			
		0 mm	to	10,000 mm		(1L + 0.1) μm	Direct measurement by a laser interferometer			
22	Gauges, special meters, measuring jigs, templates, rules	0 mm	to	330 mm		(2L + 3.5) μm	Direct measurement by a 2D microscope	KP D20		
23	Blade and surface rules	0 mm	to	2,000 mm		(5L + 2) μm	Direct measurement on a plate	KP D21		
		2,000 mm	to	3,000 mm		(5L + 12) μm	Measurement on a bed			
24*	Roughness meters	0.01 μm	to	6,000 μm		5 %	Comparative measurement by a roughness reference standard	KP DR1		
25	Roughness standards	0.01 μm	to	6,000 μm		5 %	Direct measurement by a roughness meter	KP DR1		
26	Angles from 0° to 180°	0 mm	to	7 mm	Length of the arm up to 3 m	(4.5L + 2) μm	Direct measurement by a 3D CMM	KP R2		

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Explanatory notes:

CMM – coordinate measuring machine

L – nominal length in metres

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Plane angle

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1	Levels – builder's, liquid, machine	-52 mm/m		to	52 mm/m	Division sensitivity from 0.01 mm/m	0.005mm/m	Direct measurement by a small angle generator	KP R1	
	Clinometers	-180 °		to	180 °	Division from 0.01°	0.15°			
2	Angle gauges	0 °		to	360 °		5´	Direct measurement using angle gauges	KP R2	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Mass

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1*	Non-automatic weighing instruments	0.001 g	to	2000 g		E2 class weight	$2.7 \cdot 10^{-6}$	Reference weight loading (according to OIML R111-1:2004)	KP VA1	
		2 kg	to	20 kg		F2 class weight	$1.4 \cdot 10^{-5}$			
		20 kg	to	1000 kg		M1 class weight	$5.0 \cdot 10^{-5}$			
2	Conventional weight of weights and objects							Comparison with a reference weight (according to OIML R111-1:2004) on the standard scales	KP VA2	
				1 g			0.4 mg			
		1 g	to	2 g			0.5 mg			
		2 g	to	5 g			0.6 mg			
		5 g	to	10 g			0.7 mg			
		10 g	to	20 g			0.9 mg			
		20 g	to	50 g			1.2 mg			
		50 g	to	100 g			1.9 mg			
		100 g	to	200 g			2.8 mg			
		200 g	to	500 g			5.1 mg			
		500 g	to	1 kg			10 mg			
		1 kg	to	2 kg			17 mg			
		2 kg	to	5 kg			31 mg			
		5 kg	to	10 kg			60 mg			
		10 kg	to	20 kg			90 mg			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Rotational speed

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min. unit	max. unit					
1*	Revolution meters	30 min ⁻¹	to 40,000 min ⁻¹		1.1 %	Direct measurement by a revolution generator	KP OT1	

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Force, mechanical tests

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min	unit.	max	unit.					
1	Hardness / Rockwell hardness plates and samples	70 HRA	to	85 HRA		0.40 HRA	Direct measurement	KP TV1		
		60 HRB	to	100 HRB		0.40 HRB				
	Hardness / Hardness plates									
	Shore A	0 ShA	to	100 ShA		2.0 ShA				
	Shore D	0 ShD	to	100 ShD		2.0 ShD				
	Brinell	8 HBW	to	650 HBW		1.0 %				
	Vickers	10 HV	to	2,900 HV	HV2 to HV50	1.0 %				
2*	Hardness / Hardness meters for metals						Direct measurement using reference hardness plates	KP TV1		
		Rockwell	70 HRA	to	85 HRA	0.50 HRA				
						0.50 HRB				
						0.50 HRC				
		Vickers	10 HV	to	2,000 HV	0.50 %				
		Brinell	10 HBW	to	650 HBW	0.50 %				
	Hardness / Shore hardness meters, type A,D,E,C	1 Sh	to	100 Sh		0.50 Sh				
3*	Torque / Torque wrenches	0.1 Nm	to	1,100 Nm		0.65 %	Comparative measurement by a reference torque sensor	KP S1		
		1,100 Nm	to	3,000 Nm		0.90 %				
	Torque / Torque measuring devices, torque drivers, torque sensors	0.1 Nm	to	500 Nm		0.40 %				
		500 Nm	to	2,000 Nm		1.05 %				

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min	unit.	max	unit.					
4	Moment of force / Force meters and extensometric sensors	0.001 N	to	5 kN		0.20 %	Comparative measurement by a reference force sensor	KP S2		
		5 kN	to	30 kN		0.30 %				
5*	Moment of force / Force meters and extensometric sensors	0.001 N	to	5 kN		0.20 %	Comparative measurement by a reference force sensor	KP S2		
		5 kN	to	20 kN		0.30 %				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Pressure

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1*	Deformation manometers, tyre pressure gauges, electromechanical pressure gauges (digital pressure gauges, pressure transducers with digital output of the measured quantity)							Comparative measurement by a reference digital pressure gauge	KP T1, KP T2	
		kPa	to	7 MPa	Gas	Absolute pressure	0.05% +100 Pa			
		-100 kPa	to	0 kPa	Gas	Overpressure	130 Pa			
		0 kPa	to	35 kPa			18 Pa			
35 kPa	to	7 MPa			0.05%					
		0 kPa	to	7 MPa	Liquids	Absolute pressure	0.05% +100 Pa			
		7 MPa	to	70 MPa			0.05%			
		70 MPa	to	140 MPa			0.2 %			
		0 MPa	to	1.4 MPa	Liquids	Overpressure	1,4 kPa			
		1.4 MPa	to	70 MPa			0.05 %			
		70 MPa	to	140 MPa			0.2 %			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Temperature

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1*	Direct-indicating thermometers	-30 °C	to	0 °C		0.14 °C	Comparison with a reference digital thermometer in a dry block calibrator	KP TE1		
		0 °C	to	100 °C		0.08 °C	Comparison with a reference digital thermometer in a liquid bath			
		100 °C	to	200 °C		0.22 °C	Comparison with a reference digital thermometer in a dry block calibrator			
		200 °C	to	300 °C		0.32 °C				
		300 °C	to	400 °C		0.42 °C				
		400 °C	to	500 °C		0.52 °C				
		500 °C	to	650 °C		0.67 °C				
Contactless thermometers	650 °C	to	1,100 °C		1.5 °C	Comparison with a reference digital thermometer in an air oven				
2*	Thermoelectric temperature sensors	-10 °C	to	200 °C		3.0 °C	Comparison with a reference pyrometer on target-type or cavity-type black body	KP TE4		
		200 °C	to	500 °C		6.0 °C				
		500 °C	to	800 °C		10.0 °C				
2*	Thermoelectric temperature sensors	-30 °C	to	0 °C		0.7 °C	Comparison with a reference digital thermometer in a dry block calibrator	KP TE2		
		0 °C	to	100 °C		0.7 °C				Comparison with a reference digital thermometer in a liquid bath
		100 °C	to	550 °C		0.9 °C				Comparison with a reference digital thermometer in a dry block calibrator
		550 °C	to	800 °C		2.3 °C				Comparison with a reference digital thermometer in an air oven

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
3*	Resistance temperature sensors	-30 °C	to	0 °C		0.15 °C	Comparison with a reference digital thermometer in a dry block calibrator	KP TE3	1	
		0 °C	to	100 °C		0.13 °C	Comparison with a reference digital thermometer in a liquid bath			
		100 °C	to	400 °C		0.45 °C	Comparison with a reference digital thermometer in a dry block calibrator			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Electrical quantities

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1*	DC voltage / DC voltage sources	0 mV	to	100 mV		0.0062 % + 6,1 μV	Direct measurement by a standard multimeter	KP EL2		
		0.1 V	to	1 V		0.0047 % + 16 μV				
		1 V	to	10 V		0.0047 % + 0.14 mV				
		10 V	to	100 V		0.0079 % + 2.0 mV				
		100 V	to	1000 V		0.0079 % + 20 mV				
2*	DC voltage / DC voltage meters	0 mV	to	200 mV		0.0053 % + 7.7 μV	Direct generation with a standard calibrator	KP EL1		
		0.2 V	to	2 V		0.0028 % + 15 μV				
		2 V	to	20 V		0.0028 % + 0.15 mV				
		20 V	to	200 V		0.0028 % + 1.5 mV				
		200 V	to	1000 V		0.0035 % + 17 mV				
3*	Direct current / Direct current sources	0 μA	to	10 μA		0.050 % + 6.1 nA	Direct measurement by a standard multimeter	KP EL2		
		10 μA	to	100 μA		0.074 % + 17 nA				
		0.1 mA	to	1 mA		0.075 % + 0.16 μA				
		1 mA	to	10 mA		0.034 % + 1.2 μA				
		10 mA	to	100 mA		0.034 % + 12 μA				
		100 mA	to	1 A		0.068 % + 0.15 mA				
		1 A	to	3 A		0.061 % + 1.1 mA				
4*	Direct current / Direct current meters	0 μA	to	200 μA		0.020 % + 73 nA	Direct generation with a standard calibrator	KP EL1		
		0.2 mA	to	2 mA		0.016 % + 0.13 μA				
		2 mA	to	20 mA		0.009 % + 0.94 μA				
		20 mA	to	200 mA		0.012 % + 9.7 μA				
		0.2 A	to	2 A		0.018 % + 0.13 mA				

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location		
		min.	unit	max.	unit							
		2 A	to	30 A		0.069 % + 2.5 mA						
		30 A	to	1500 A		0.42 % + 0.13 A	Simulation using current coil					
5*	AC voltage / AC voltage sources	0.1 mV	to	100 mV	10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.086 % + 36 μV 0.16 % + 59 μV 0.69 % + 93 μV	Direct measurement by a standard multimeter	KP EL2				
		0.1 V	to	1 V	10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.076 % + 0.43 mV 0.14 % + 0.73 mV 0.71 % + 0.81 mV						
		1 V	to	10 V	10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.076 % + 4.4 mV 0.14 % + 7.3 mV 0.71 % + 17 mV						
		10 V	to	100 V	10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.076 % + 44 mV 0.14 % + 73 mV 0.71 % + 81 mV						
		100 V	to	750 V	10 Hz to 20 kHz 20 kHz to 50 kHz 50 kHz to 100 kHz	0.07 % + 0.35 V 0.14 % + 0.59 V 0.61 % + 1.5 V						
6*	AC voltage / AC voltage meters	0.1 mV	to	200 mV	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz 20 kHz to 50 kHz	0.16 % + 64 μV 0.049 % + 69 μV 0.12 % + 86 μV 0.20 % + 98 μV				Direct generation with a standard calibrator	KP EL1	
		0.2 V	to	2 V	10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz	0.082 % + 0.38 mV 0.042 % + 0.29 mV 0.12 % + 0.46 mV						

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
		2 V	to	20 V	20 kHz to 50 kHz 10 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz	0.17 % + 0.73 mV 0.084 % + 3.8 mV 0.042 % + 2.7 mV 0.12 % + 4.7 mV				
		20 V	to	200 V	20 kHz to 50 kHz 30 Hz to 45 Hz 45 Hz to 10 kHz 10 kHz to 20 kHz	0.17 % + 5.6 mV 0.082 % + 31 mV 0.10 % + 43 mV 0.14 % + 56 mV				
		200 V	to	1000 V	20 kHz to 40 kHz 30 Hz to 45 Hz 45 Hz to 1 kHz 1 kHz to 5 kHz 5 kHz to 10 kHz	0.17 % + 61 mV 0.082 % + 0.45 V 0.09 % + 0.28 V 0.14 % + 0.38 V 0.16 % + 0.49 V				
7*	Alternating current / Alternating current sources	0,1 µA	to	100 µA	10 Hz to 1 kHz	0.15 % + 7.3 nA	Direct measurement by a standard multimeter	KP EL2		
		0,1 mA	to	1 mA	10 Hz to 5 kHz	0.10 % + 0.51 µA				
		1 mA	to	10 mA	10 Hz to 5 kHz	0.10 % + 5.1 µA				
		10 mA	to	100 mA	10 Hz to 5 kHz	0.10 % + 51 µA				
		0,1 A	to	1 A	10 Hz to 5 kHz	0.10 % + 0.51 mA				
		1 A	to	3 A	10 Hz to 5 kHz	0.15 % + 3.4 mA				
8*	Alternating current / Alternating current meters	20 µA	to	200 µA	10 Hz to 45 Hz 45 Hz to 1 kHz	0.25 % + 0.40 µA 0.08 % + 0.30 µA	Direct generation with a standard calibrator	KP EL1		
		0,2 mA	to	2 mA	10 Hz to 45 Hz 45 Hz to 1 kHz	0.23 % + 0.40 µA 0.06 % + 0.30 µA				

The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
		2 mA	to	20 mA		10 Hz to 45 Hz	0.23 % + 4.5 μA			
		20 mA	to	200 mA		45 Hz to 1 kHz	0.05 % + 3.0 μA			
		0.2 A	to	2 A		10 Hz to 45 Hz	0.23 % + 45 μA			
		2 A	to	30 A		45 Hz to 1 kHz	0.05 % + 30 μA			
						10 Hz to 45 Hz	0.23 % + 0.45 mA			
						45 Hz to 1 kHz	0.06 % + 0.30 mA			
						30 Hz to 45 Hz	0.23 % + 4.5 mA			
						45 Hz to 100 Hz	0.06 % + 3.0 mA			
						100 Hz to 1 kHz	0.60 % + 5.8 mA			
		30 A	to	1500 A		30 Hz to 60 Hz	0.42 % + 0.13 A	Simulation using current coil		
9*	DC resistance / DC resistance / DC resistance meters							Direct generation with a standard calibrator	KP EL 1	
							8.7 mΩ			
							1 Ω			
							8.9 mΩ			
							10 Ω			
							11 mΩ			
							100 Ω			
							20 mΩ			
							1 kΩ			
							0.16 Ω			
							1.6 Ω			
							10 kΩ			
							15 Ω			
							100 kΩ			
							0.29 kΩ			
							1 MΩ			
							8.0 kΩ			
							10 MΩ			
							0.98 MΩ			
							100 MΩ			
							22 MΩ			
		0 Ω	to	100 Ω			0.018 % + 88 m Ω			
		100 Ω	to	330 Ω			0.012 % + 0.11 Ω			
		330 Ω	to	1 kΩ			0.01 % + 0.19 Ω			

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2024 of 14/02/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
		1 kΩ	to	3.3 kΩ		0.012 % + 0.29 Ω				
		3,3 kΩ	to	10 kΩ		0.0082 % + 1.4 Ω				
		10 kΩ	to	33 kΩ		0.012 % + 2.2 Ω				
		33 kΩ	to	100 kΩ		0.0066 % + 18 Ω				
		100 kΩ	to	330 kΩ		0.011 % + 24 Ω				
		330 kΩ	to	1 MΩ		0.0066 % + 0.18 kΩ				
		1 MΩ	to	3.3 MΩ		0.013 % + 0.24 kΩ				
		3.3 MΩ	to	10 MΩ		0.0075 % + 1.8 kΩ				
		10 MΩ	to	33 MΩ		0.052 % + 6.4 kΩ				
		33 MΩ	to	100 MΩ		0.064 % + 0.21 MΩ				
		100 MΩ	to	330 MΩ		1.3 % + 1.9 MΩ				
		330 MΩ	to	1 GΩ		2.4 % + 13 MΩ				
10*	Capacitance / Electrical Capacitance Meters			1 nF	1 kHz	0.48 %	Direct generation with a standard calibrator	KP EL1		
				2 nF	1 kHz	0.73 %				
				5 nF	1 kHz	0.70 %				
				10 nF	1 kHz	0.54 %				
				100 nF	1 kHz	0.41 %				
				1 μF	1 kHz	0.65 %				
				10 μF	1 kHz	0.98 %				
		10 μF	to	100 μF		1.2 %	Comparison with a standard multimeter			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
CAB number 2301, Calibration Laboratory
Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Time quantities and frequency

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range		Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min. unit	max. unit					
1	Time interval / time meters, stopwatches, timers	1 s	to 86,400 s		0.5 s	Comparison with a reference digital stopwatch	KP Č1	
2*	Frequency / low frequency and high frequency counters	1 Hz	to 100 kHz		1.7.10 ⁻⁶	Direct generation by reference calibrator	KP EL1	
		100 KHz	to 1,000 KHz		2.3.10 ⁻⁶			
			10 MHz		1.7.10 ⁻⁵			

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

M & B Calibr, spol. s r.o.
 CAB number 2301, Calibration Laboratory
 Krumlovská 1454/26, 664 91 Ivančice

CMC for the field of measured quantity: Humidity

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Location
		min.	unit	max.	unit					
1*	Relative humidity meters except psychrometers	10 % RH	to	95 % RH	(20 to 40) °C	2.3% RH	Comparative measurement by a reference hygrometer in a humidity generator	KP VL1		

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself. "