

**The Appendix is an integral part of
Certificate of Accreditation No. 369/2024 of 29/07/2024**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

Recovera Facility a.s.
CAB number 1767, Immission Measurement Laboratory
Drčkova 2798/7, 628 00 Brno

Tests:

Ordinal number ¹	Test procedure/method name	Test procedure/method identification ²	Tested subject	Degrees of freedom ³
1*	Determination of the mass concentration of suspended particles PM ₁ , PM _{2,5} , PM ₁₀ by optoelectronic method	SOP 01 (Grimm manual; Palas manual)	Ambient air (immissions)	-
2*	Determination of the mass concentration of sulphur dioxide (SO ₂) by UV fluorescence method	SOP 02 (ČSN EN 14212)	Ambient air (immissions)	-
3*	Determination of the mass concentration of nitrogen oxides (NO, NO ₂ , NO _x) by chemiluminescence method	SOP 03 (ČSN EN 14211)	Ambient air (immissions)	-
4*	Determination of the mass concentration of carbon monoxide (CO) by chemiluminescence method	SOP 04 (ČSN EN 14626)	Ambient air (immissions)	-
5*	Determination of the mass concentration of ozone (O ₃) by UV absorption photometry method	SOP 05 (ČSN EN 14625)	Ambient air (immissions)	-

¹ asterisk at the ordinal number identifies the tests, which the laboratory is qualified to carry out outside the permanent laboratory premises

² if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)

³ the laboratory does not apply a flexible approach to the scope of accreditation

"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself."