Accredited entity according to ČSN EN ISO/IEC 17025:2018:

DEOM s.r.o.

CAB number 2407, Calibration Laboratory Jinonická 804/80, Košíře, 158 00 Praha 5

CMC for the field of measured quantity: Length

Ord. number	Calibrated quantity / Subject of calibration		ominal	range	Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification	Location
		min uni	t	max unit					
1*	Coordinate measuring machine – optical							KP1	
	axis X and Y	0 mm	to	2,500 mm		(L/660 +0.6) μm	Measuring with a glass ruler		
	Z-axis						Measurement using parallel gauge		
		0 mm	to	400 mm		$(L/420 +1.0) \mu m$	blocks		
2*	Multisensor equipment of coordinate measuring machine – optical							KP2	
							Measurement with a calibration		
	Touch probe	0 mm	to	12 mm		0.4 μm	sphere		
							Measurement using a parallel gauge		
	Optical distance sensor	0 mm	to	24 mm		0.1 μm	block		

Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

Explanatory notes:

L nominal length [mm]

The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

[&]quot;This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself."