

**The Appendix is an integral part of  
Certificate of Accreditation No: 409/2024 of 20/08/2024**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**České vysoké učení technické v Praze**  
CAB number 1659, Laboratory for the Photovoltaic Systems Diagnostics (LPVSD)  
Technická 2, 166 27 Praha 6 – Dejvice

*The laboratory provides opinions and interpretations of the test results.*

**Tests:**

Ordinal number <sup>1</sup>	Test procedure / method name	Test procedure / method identification <sup>2</sup>	Tested subject	Degrees of freedom <sup>3</sup>
1	Measurement of Volt-Ampere Characteristic	BZP_01_A (ČSN EN IEC 60904-1 ed. 3, cl. 7; ČSN EN 60904-1 ed. 2:2007, cl. 7; ČSN EN IEC 61215-2 ed. 2, cl. 4.2, 4.6, 4.7; ČSN EN 61215-2:2017, cl. 4.2, 4.6, 4.7; ČSN EN 61215:1997, cl. 10.2, 10.6, 10.7; ČSN EN 61215 ed. 2:2006, cl. 10.2, 10.6, 10.7; ČSN EN 61646:1999, cl. 10.2, 10.6, 10.7; ČSN EN 61646 ed. 2:2009, cl. 10.2, 10.6, 10.7)	Photovoltaic Modules	-
2	Dielectric withstand test	BZP_03_A (ČSN EN IEC 61730-2 ed. 2, cl. 10.13; ČSN EN IEC 61215-2 ed. 2, cl. 4.3; ČSN EN 61215-2:2017, cl. 4.3; ČSN EN 61215:1997, cl. 10.3; ČSN EN 61215 ed. 2:2006, cl. 10.3; ČSN EN 61646:1999, cl. 10.3; ČSN EN 61646 ed. 2:2009, cl. 10.3)	Photovoltaic Modules	-
3	Insulation test by applied voltage below 500 V	BZP_04_A (ČSN EN IEC 61215-2 ed. 2, cl. 4.3; ČSN EN 61215-2:2017, cl. 4.3; ČSN EN 61215:1997, cl. 10.3; ČSN EN 61215 ed. 2:2006, cl. 10.3; ČSN EN 61646:1999, cl. 10.3; ČSN EN 61646 ed. 2:2009, cl. 10.3)	Photovoltaic Modules	-
4	Continuity test of equipotential bonding by applied current below 30 A	BZP_05_A (ČSN EN IEC 61730-2 ed. 2, cl. 10.11)	Photovoltaic Modules	-

**The Appendix is an integral part of  
Certificate of Accreditation No: 409/2024 of 20/08/2024**

**Accredited entity according to ČSN EN ISO/IEC 17025:2018:**

**České vysoké učení technické v Praze**  
CAB number 1659, Laboratory for the Photovoltaic Systems Diagnostics (LPVSD)  
Technická 2, 166 27 Praha 6 – Dejvice

Ordinal number <sup>1</sup>	Test procedure / method name	Test procedure / method identification <sup>2</sup>	Tested subject	Degrees of freedom <sup>3</sup>
5*	Visual Inspection	BZP_06_A (ČSN EN IEC 61215-2 ed. 2, cl. 4.1; ČSN EN 61215-2:2017, cl. 4.1; ČSN EN 61215:1997, cl. 10.1; ČSN EN 61215 ed. 2:2006, cl. 10.1; ČSN EN 61646:1999, cl. 10.1; ČSN EN 61646 ed. 2:2009, cl. 10.1)	Photovoltaic Modules	-

<sup>1</sup> asterisk at the ordinal number identifies the tests, which the laboratory is qualified to carry out outside the permanent laboratory premises

<sup>2</sup> if the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest valid edition of the specified procedure is used (including any changes)

<sup>3</sup> the laboratory does not apply a flexible approach to the scope of accreditation

**Explanations:**

BZP\_XX\_A – LDFS Test Procedure

---

*"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and the certificate itself. "*