

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
 CAB number 2362, Calibration laboratory  
 Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Pressure**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Deformation and digital manometers, pressure transducers and pressure measuring chains	-95 kPa	to	-7 kPa		relative pressure gas	0.04 %	Comparison with a standard calibrator	KL-PM-0101	
		-7 kPa	to	14 kPa			0.0028 kPa			
		14 kPa	to	14 MPa		relative pressure liquid	0.02 %	Comparison with a piston manometer		
		14 MPa	to	60 MPa			0.03 %	Comparison with a digital manometer		
60 MPa	to	70 MPa		absolute pressure gas	0.1 %	Comparison with a standard calibrator				
5 kPa	to	70 kPa			0.028 kPa	Comparison with a piston manometer				
70 kPa	to	14 MPa		absolute pressure liquid	0.02 % + 0.014 kPa	Comparison with a digital manometer				
14 MPa	to	60 MPa			0.03 % + 0.1 kPa	Comparison with a piston manometer				
60 MPa	to	70 MPa			0.10 % + 0.1 kPa	Comparison with a digital manometer				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
 CAB number 2362, Calibration laboratory  
 Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Temperature**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
1*	Resistance temperature sensors (with/without a transducer), direct indicating thermometers and measuring chains with resistance temperature sensors									
				-196 °C	to	-40 °C		0.15 °C	Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0001
				-40 °C	to	0 °C		0.20 °C		
						0 °C		0.05 °C		
				0 °C	to	100 °C		0.04 °C		
				100 °C	to	200 °C		0.05 °C		
				200 °C	to	300 °C		0.06 °C		
				300 °C	to	420 °C		0.15 °C		
				420 °C	to	660 °C		0.18 °C		
2*	Thermoelectric temperature sensors (with/without a transducer), direct indicating thermometers and measuring chains with thermocouples								Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0002
				-196 °C	to	-40 °C		0.6 °C		
				-40 °C	to	200 °C		0.3 °C		
				200 °C	to	400 °C		0.6 °C		
				400 °C	to	660 °C		0.9 °C		
				400 °C	to	900 °C		0.9 °C	Comparison with standard thermocouples S, B, Pt-Pd in horizontal ovens	
				900 °C	to	1,100 °C		1.0 °C		
				1,100 °C	to	1,200 °C		1.5 °C		
				1,200 °C	to	1,400 °C		2.0 °C		
				1,400 °C	to	1,553 °C		2.8 °C		

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
 CAB number 2362, Calibration laboratory  
 Raisova 547, Holínské Předměstí, 506 01 Jičín

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min.	unit	max.	unit					
3*	Dial thermometers	-40 °C	to	200 °C		0.2 °C	Comparison with a Pt100 standard in liquid baths and dry block calibrators	KL-PM-0005		
		200 °C	to	500 °C		0.7 °C				
		500 °C	to	660 °C		1.2 °C				

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
 CAB number 2362, Calibration laboratory  
 Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Air humidity**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min	unit	max	unit					
1*	Relative humidity / hygrometers and measuring chains incl. humidity probes	5 % RH	to	30 % RH		Air Temperature (7 to 60) °C	1.2 % RH	Comparison with a reference hygrometer	KL-PM-0201	
		30 % RH	to	50 % RH			1.3 % RH			
		50 % RH	to	70 % RH			1.4 % RH			
		70 % RH	to	80 % RH			1.5 % RH			
		80 % RH	to	90 % RH			1.6 % RH			
		90 % RH	to	95 % RH			1.8 % RH			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

**JSP, s.r.o.**  
 CAB number 2362, Calibration laboratory  
 Raisova 547, Holínské Předměstí, 506 01 Jičín

**CMC for the field of measured quantity: Electrical quantities**

Ord. number <sup>1</sup>	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty <sup>2</sup>	Calibration principle	Calibration procedure identification <sup>3</sup>	Location
		min	unit	max	unit					
1*	Measurement and simulation of temperature sensor signals (resistance temperature sensors, thermocouple temperature sensors)	0 Ω	to	600 Ω		0.007 % + 3 mΩ	Comparison with a reference multimeter	KL-PM-0006		
		600 Ω	to	6,000 Ω		0.007 % + 30 mΩ				
		-10 mV	to	100 mV		0.004 % + 1.7 μV				
	Measurement and simulation of unified output signals	0 V	to	10 V		0.0035 % + 47 μV	Comparison with a reference multimeter			
		0 mA	to	20 mA		0.005 % + 0.0008 mA	Indirect current measurement			

<sup>1</sup> Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

<sup>2</sup> The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

<sup>3</sup> If the document identifying the calibration procedure is dated only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

*"This document is an appendix to the certificate of accreditation. In case of any discrepancies between the English and Czech versions, the Czech version shall prevail, both for the certificate appendix and for the certificate itself."*